

Low Detection Rate of *Bordetella pertussis* Using the BioFire FilmArray Respiratory Panel 2plus

Sophie Guillot, Assaf Mizrahi, Nathalie Armatys, Laureen Chat, Alban Le Monnier, Sylvain Brisse, Julie Toubiana

► **To cite this version:**

Sophie Guillot, Assaf Mizrahi, Nathalie Armatys, Laureen Chat, Alban Le Monnier, et al.. Low Detection Rate of *Bordetella pertussis* Using the BioFire FilmArray Respiratory Panel 2plus. *Open Forum Infectious Diseases*, Oxford University Press, 2020, 7 (8), pp.2169 - 2171. 10.1093/ofid/ofaa267. pasteur-03255758

HAL Id: pasteur-03255758

<https://hal-pasteur.archives-ouvertes.fr/pasteur-03255758>

Submitted on 9 Jun 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Low Detection Rate of *Bordetella pertussis* Using the BioFire FilmArray Respiratory Panel 2plus

Sophie Guillot,^{1,2} Assaf Mizrahi,^{3,4} Nathalie Armatys,^{1,2} Laureen Chat,³ Alban Le Monnier,^{3,4} Sylvain Brisse,^{1,2} and Julie Toubiana^{1,2,5,6}

¹Biodiversity and Epidemiology of Bacterial Pathogens, Institut Pasteur, Paris, France, ²National Reference Center for Whooping Cough and Other *Bordetella* Infections, Institut Pasteur, Paris, France, ³Clinical Microbiology Laboratory, Saint Joseph Hospital, Paris, France, ⁴Institut Micalis, UMR1319, Université Paris-Saclay, INRAE, AgroParisTech, Bactéries Pathogènes et Santé, Chatenay-Malabry, France, and ⁵General Paediatrics and Paediatric Infectious Diseases Department, Hôpital Necker-Enfants Malades, APHP, Université de Paris, Paris, France

Syndromic respiratory panels are increasingly used worldwide. Their performance for detection of *Bordetella pertussis* needs to be evaluated. We found that the FilmArray Respiratory Panel 2plus (RP2+) assay, which uses the pertussis toxin promoter target for *B. pertussis*, can only detect highly charged samples. Negative RP2+ results should not be interpreted as an absence of *B. pertussis* in clinical samples.

Keywords. *Bordetella pertussis*; FilmArray respiratory panel; molecular diagnosis; whooping cough.

Bordetella pertussis, the main agent of whooping cough (also called pertussis), continues to cause infections despite high levels of vaccination and has even re-emerged in several countries [1]. Clinical diagnosis of whooping cough is difficult given the nonspecificity of symptoms in adults and vaccinated individuals [2]. Co-infections with viruses (eg, respiratory syncytial virus) may also confound a diagnosis of pertussis in infants, especially if these are the only agents tested for [3]. The laboratory confirmation of whooping cough is therefore essential for accurate diagnosis and treatment of the index case, as well as for surveillance purposes and prevention of disease transmission. Laboratory confirmation mostly relies on direct detection methods on a nasopharyngeal (NP) specimen, such as culture and quantitative polymerase chain reaction (PCR). PCR is the most sensitive approach when performed within the 3 first weeks after disease onset, but molecular *B. pertussis* detection methods suffer from a trade-off between sensitivity

and specificity [4]. The target with the highest analytical sensitivity is the insertion sequence IS481, which is repeated >240 times in the *B. pertussis* genome [5]. However, IS481 is also present in *Bordetella holmesii* (range, 28–36 copies) [6], which can also be detected in respiratory samples from patients with whooping cough–like illness [7], as well as in some *Bordetella bronchiseptica* isolates (≤ 1 copy) [6]. *B. pertussis*–specific PCRs targeting the pertussis toxin promoter region *ptxP* have been described, but these have reduced sensitivity compared with IS481, as only a single copy is present in the genome [4].

The FilmArray Respiratory Panel 2 plus (RP2+) is a multiplex in vitro diagnostic test for the simultaneous and rapid detection of 22 respiratory pathogens (18 viruses and 4 bacteria species including *B. pertussis*) directly from NP samples. It consists of automated nucleic acid extraction, nested multiplex PCR, and analysis of generated end point melting curve data. This multiplex panel presents the advantages of speed and could lead to pertussis diagnosis when the symptoms are not typical. The RP2+ as the previous RP genetic target for *B. pertussis* is *ptxP* [8]. In a recent multicenter evaluation of the FilmArray RP2 in nasopharyngeal swab samples, 2 out of 3 positive *B. pertussis* samples (according to the comparator RP) were detected [9].

The present study aimed to evaluate the performance of the FilmArray RP2+ for detection of *B. pertussis* on clinical samples when compared with a simplex quantitative PCR (qPCR) using the same target, with the IS481 qPCR considered the reference standard.

METHODS

Sample Selection Based on *B. pertussis* Detection With an In-house *ptxP* qPCR

An in-house control *ptxP* real-time PCR targeting the promoter region of the pertussis toxin gene (named *ptxA*–Pr) was used as the comparator. This assay (*ptxP* qPCR) was previously shown to have a lower limit of detection (LLOD) of 30 CFU per PCR reaction [7]. The selected samples were clinical nasopharyngeal aspirates/swabs from patients with pertussis symptoms ($n = 13$) and were completed with quality control clinical samples (Quality Control for Molecular Diagnostics: www.qcmd.org; $n = 5$), all stored at -80°C at the French National Reference Center of Whooping Cough. All the selected samples were previously analyzed by qPCR and found positive for IS481 and negative for hIS1001 (specific for *B. holmesii* detection). The samples were collected continuously between 2018 and 2019 and tested using the in-house *ptxP* qPCR. Among these samples, we selected 16 to represent a range of positive *ptxP* threshold cycle (Ct) values and 2 samples for being below the detection limit of *ptxP* qPCR (no Ct value with *ptxP*) while

Received 26 April 2020; editorial decision 22 June 2020; accepted 26 June 2020.

Correspondence: J. Toubiana, MD, PhD, Université de Paris, Biodiversity and Epidemiology of Bacterial Pathogens, Institut Pasteur, 28 rue du Docteur Roux, F-75724 Paris, France (julie.toubiana@pasteur.fr).

Open Forum Infectious Diseases®

© The Author(s) 2020. Published by Oxford University Press on behalf of Infectious Diseases Society of America. This is an Open Access article distributed under the terms of the Creative Commons Attribution-NonCommercial-NoDerivs licence (<http://creativecommons.org/licenses/by-nc-nd/4.0/>), which permits non-commercial reproduction and distribution of the work, in any medium, provided the original work is not altered or transformed in any way, and that the work is properly cited. For commercial re-use, please contact journals.permissions@oup.com
DOI: 10.1093/ofid/ofaa267

being positive on *IS481* qPCR, with Ct values of 34 and 36 for this higher-copy target. The selected samples were coded with a letter designation to be tested blindly with FilmArray RP2+.

All French bacteriological samples were collected, coded, shipped, managed, and analyzed according to the National Reference Center protocols and received approval from the French Supervisory Ethics Authority (CNIL, No. 1474593).

FilmArray RP2+ Analyses

Samples were tested using the FilmArray RP2+ according to the manufacturer's instructions [8]. Briefly, for each sample, 1.0 mL of hydration solution (water) was added to the pouch to rehydrate the reagents. In total, 300 µL of the reserve sample was added to 500 µL of sample buffer mix and thoroughly combined. Then, 300 µL of sample/sample buffer mix was added to the pouch, which was then loaded onto the instrument (system 2.0). Each run contained internal process controls for extraction, dilution, and PCR. The FilmArray test consists of automated nucleic acid extraction, nested multiplex PCR, and analysis of generated end point melting curve data. This process took approximately 1 hour per run. Valid run results were reported as "detected" or "not detected" for each target.

Performance Analysis

To compare the performance of RP2+ and in-house *ptxP* qPCR, all samples were analyzed in parallel by *IS481* qPCR (Argene *Bordetella* R-gene, Biomerieux). The detection rate was estimated based on negative or positive test results of RP2+ and

in-house *ptxP* qPCR compared with *IS481*, which is considered the reference standard.

RESULTS

The results are presented in Figure 1. Twelve samples were positive for *B. pertussis* using both the FilmArray RP2+ and the in-house *ptxP* qPCR. The range of *ptxP* Ct values for these samples was 18–35. In contrast, 4 samples were found to be negative using the FilmArray RP2+ assay; all had a Ct value >35 using the in-house *ptxP* qPCR. These 4 discrepant samples were retested by the in-house *ptxP* qPCR, starting from the clinical material, thus including extraction and amplification, and were confirmed positive with Ct values >35 (37–40). Finally, the 2 samples that were negative with the in-house *ptxP* assay but positive using the qPCR *IS481* with a Ct value of 34 and 36, respectively, were also negative with RP2+ (Figure 1). Therefore, while the detection rate for *B. pertussis* of in-house *ptxP* qPCR was fixed at 89% in our study, the detection rate of RP2+ was 67%, revealing a lower detection rate for RP2+, when considering qPCR *IS481* the reference standard.

DISCUSSION

The FilmArray respiratory multiplex panel is being increasingly used for diagnosis of respiratory infections. It presents the advantages of multipathogen testing, allowing for detection of nonanticipated pathogens using broad diagnosis coverage. However, multiplexing often comes at the expense of analytical




Figure 1. Comparison of Ct values obtained with the FilmArray RP2+ and *ptxP* qPCR. Samples (n = 18, A to R) were ordered by increasing Ct value of the *ptxP* qPCR. The Ct values of the reference standard *IS481* qPCR are indicated by black triangles. Green, red, and purple circles denote samples that were positive for both the *ptxP* qPCR and RP2+ methods (green), *ptxP* qPCR only (red), or none (purple). Abbreviation: qPCR, quantitative polymerase chain reaction.

sensitivity [10]. This issue is critical for detection of *B. pertussis*, which can be present at low levels in clinical samples, and the abundance of which quickly decreases over time despite symptoms being present [10, 11]. Previous studies have compared the performance of the first version of the FilmArray RP to qPCR targeting the insertion sequence IS481, which is the most frequently used target for the molecular diagnosis of pertussis. Two of these studies did not find a loss of analytical sensitivity for *B. pertussis* when using the FilmArray RP, but they were designed to assess the entire panel and *B. pertussis* was therefore not their primary focus [12, 13]. Furthermore, in the Pierce et al. study, all samples detected positive by FilmArray RP had low Ct values (<30) with real-time PCR targeting IS481 [13]. Another study compared the performance of the FilmArray RP with the Focus qPCR assay (Focus Diagnostic, Murrieta, CA, USA), which targets IS481. Seventy-one specimens from patients, which had been tested positive for *B. pertussis*, were analyzed and compared. The authors concluded that the FilmArray RP assay detects ~30% less cases of *B. pertussis* than the Focus assay [14]. In a recent multicenter evaluation of the FilmArray RP2 in nasopharyngeal swab samples, only 3 positive *B. pertussis* samples were tested and compared with the previous FilmArray RP [9].

Here, we evaluated the performance of the FilmArray RP2+ to detect *B. pertussis* using a comparative design based on an in-house simplex *ptxP* target assay. The use of archived frozen clinical samples allowed us to select clinical samples representing a wide range of abundances of *B. pertussis*. We selected a range of positive samples based on *ptxP*, a single copy target that is known to be very specific, with the advantage of discriminating between *B. pertussis* and other *Bordetella* spp., *B. holmesii* in particular [7], but less sensitive than the multicopy golden standard IS481 target for the molecular diagnosis of whooping cough.

As observed in our report, the samples with a *ptxP* qPCR Ct value >35 (corresponding to IS481 >28) were not detected with the FilmArray RP2+. Hence, although the FilmArray RP2+ shares the advantage of higher specificity than the standard qPCR targeting IS481, it appears less sensitive than the simplex *ptxP* qPCR assay and a fortiori than the reference standard. Depending on the age of the patient, vaccination status, duration of the cough before biological sampling, and sample quality, a value of Ct IS481 >28 is not infrequent [15, 16]. In 2019, we received at the French National Reference Center of whooping cough 50 clinical samples, one-third of which had a *ptxP* Ct value >35, implying that one-third of samples would have been missed using RP2+ if it would not have been complemented by qPCR IS481. Most of these samples were from previously vaccinated children or adults, who are more likely to present with atypical symptoms inducing a delay between symptom onset and sampling >3 weeks. This could explain a low bacterial load. Overall, although the number

of tested samples is limited, this work calls for caution when interpreting Bp-negative results when the FilmArray RP2+ is used. Depending on the clinical and epidemiological contexts, it may be important to control the negativity of samples using more sensitive, dedicated *B. pertussis* diagnostic methods. Besides, other additional qPCR with good sensitivity targeting sequences such as *hIS1001* can be used to discriminate between *B. pertussis* and *B. holmesii*, which is also involved in pertussis-like symptoms.

Acknowledgments

Financial support. This work was supported financially by Santé Publique France (Saint-Maurice, France) and by continuous institutional support from Institut Pasteur.

Potential conflicts of interest. All authors: no reported conflicts of interest. All authors have submitted the ICMJE Form for Disclosure of Potential Conflicts of Interest. Conflicts that the editors consider relevant to the content of the manuscript have been disclosed.

Author contributions. S.G., S.B., and J.T. contributed to the conception and design of the study. S.G., A.M., N.A., and L.C. contributed to the acquisition of data. J.T., S.G., and A.L. analyzed and interpreted the data. S.G. and J.T. wrote the first draft of the article. J.T. and S.B. supervised the study. All authors revised the manuscript for important intellectual content and approved the final version submitted for publication.

References

1. Stefanelli P. Pertussis: identification, prevention and control. *Adv Exp Med Biol* **2019**; 1183:127–36.
2. Crowcroft NS, Pebody RG. Recent developments in pertussis. *Lancet* **2006**; 367:1926–36.
3. Nuolivirta K, Koponen P, He Q, et al. *Bordetella pertussis* infection is common in nonvaccinated infants admitted for bronchiolitis. *Pediatr Infect Dis J* **2010**; 29:1013–5.
4. Wirsing von König CH. Pertussis diagnostics: overview and impact of immunization. *Expert Rev Vaccines* **2014**; 13:1167–74.
5. Weigand MR, Peng Y, Loparev V, et al. The history of *Bordetella pertussis* genome evolution includes structural rearrangement. *J Bacteriol* **2017**; 199:e00806–16.
6. Tizolova A, Guiso N, Guillot S. Insertion sequences shared by *Bordetella* species and implications for the biological diagnosis of pertussis syndrome. *Eur J Clin Microbiol Infect Dis* **2013**; 32:89–96.
7. Njamkepo E, Bonacorsi S, Debruyne M, et al. Significant finding of *Bordetella holmesii* DNA in nasopharyngeal samples from French patients with suspected pertussis. *J Clin Microbiol* **2011**; 49:4347–8.
8. BioFire Diagnostics. FilmArray Respiratory Panel 2 (RP2plus) Instruction Booklet RFIT-ASY-0136. Salt Lake City, UT: BioFire Diagnostics LLC; **2017**.
9. Leber AL, Everhart K, Daly JA, et al. Multicenter evaluation of BioFire FilmArray respiratory panel 2 for detection of viruses and bacteria in nasopharyngeal swab samples. *J Clin Microbiol* **2018**; 56:e01945–17.
10. van der Zee A, Schellekens JF, Mooi FR. Laboratory diagnosis of pertussis. *Clin Microbiol Rev* **2015**; 28:1005–26.
11. Nakamura Y, Kamachi K, Toyoizumi-Ajisaka H, et al. Marked difference between adults and children in *Bordetella pertussis* DNA load in nasopharyngeal swabs. *Clin Microbiol Infect* **2011**; 17:365–70.
12. Couturier MR, Barney T, Alger G, et al. Evaluation of the FilmArray® respiratory panel for clinical use in a large children's hospital. *J Clin Lab Anal* **2013**; 27:148–54.
13. Pierce VM, Elkan M, Leet M, et al. Comparison of the Idaho Technology FilmArray system to real-time PCR for detection of respiratory pathogens in children. *J Clin Microbiol* **2012**; 50:364–71.
14. Jerris RC, Williams SR, MacDonald HJ, et al. Testing implications of varying targets for *Bordetella pertussis*: comparison of the FilmArray respiratory panel and the focus *B. pertussis* PCR assay. *J Clin Pathol* **2015**; 68:394–6.
15. Waters V, Jamieson F, Richardson SE, et al. Outbreak of atypical pertussis detected by polymerase chain reaction in immunized preschool-aged children. *Pediatr Infect Dis J* **2009**; 28:582–7.
16. Tatti KM, Wu KH, Tondella ML, et al. Development and evaluation of dual-target real-time polymerase chain reaction assays to detect *Bordetella* spp. *Diagn Microbiol Infect Dis* **2008**; 61:264–72.